

Certificate of Analysis List

For request number 1270583

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
1440053	9135	N/A	N/A	Sodium Chloride Standard Solution, 491

Total Enclosures: 1

Confidencial
GRUENTEC CÍA. LTDA.



An ISO 9001 Certified Company

Certificate of Analysis

Page 1

COMMODITY: **Sodium Chloride Standard Solution, 491**COMMODITY NUMBER: **1440053**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A9135****5/20/2019****5/20/2019**

TEST	SPECIFICATIONS	RESULTS
Conductivity at 25 °C	990 to 1010 uS/cm	998.0 uS/cm

The expiration date is May 2024

The item 1440053 is traceable to NIST standards SRM 2201 Sodium Chloride LOT N/A.

Confidencial
GRUENTEC CÍA. LTDA.

A handwritten signature in cursive script that reads "Scott Als".

Certified by _____

Scott Als
Analytical Services Chemist